

CONTENTS

	Page
ACKNOWLEDGEMENTS	iii
ABSTRACT (ENGLISH)	iv
ABSTRACT (THAI)	vi
LIST OF TABLES	xi
LIST OF FIGURES	xiii
ABBREVIATIONS AND SYMBOLS	xviii
CHAPTER 1 INTRODUCTION	1
1.1 Introduction of Sequential Injection Analysis, SIA	1
1.2 System configuration	3
1.3 Application of SIA system	4
1.4 Sequential injection analyzer	7
1.5 Ascorbic acid	12
1.5.1 Determination method of ascorbic acid	14
1.5.2 Spectrophotometric methods	15
1.6 Acetic acid in Vinegars	16
1.6.1 Analysis of vinegars	17
1.7 Aims of the Thesis	18
CHAPTER 2 EXPERIMENTAL, RESULTS AND DISCUSSION	19
2.1 Instruments, apparatus and chemicals	19
2.1.1 Instruments and apparatus	19
2.1.2 Chemicals	19
2.2 Samples	20
2.2.1 Sample for ascorbic acid determination	20
2.2.2 Sample for acetic acid determination	20
2.3 Preparation of solution	21
2.3.1 Preparation of reagents used in SIA system	21

	Page
2.3.2 Preparation of reagent used in the standard method	22
2.4 Determination of ascorbic acid by sequential injection spectrophotometric method	23
2.4.1 Schematic diagram	23
2.4.2 Procedure	24
2.4.3 Optimization for conditions for sequential injection determination of ascorbic acid	25
2.4.3.1 Effect of acidic potassium permanganate concentration	25
2.4.3.2 Effect of acidic potassium permanganate aspiration volume	27
2.4.3.3 Effect of ascorbic acid aspiration volume	30
2.4.3.4 Effect of pump speed (flow rate)	32
2.4.3.5 Summary of the conditions used	33
2.4.3.6 Calibration range	34
2.4.3.7 Precision	36
2.4.3.8 Determination of ascorbic acid in vitamin C tablet sample	38
2.4.3.9 Recovery study	41
2.5 Determination of acetic acid by sequential injection spectrophotometric method	45
2.5.1 Schematic Diagram	45
2.5.2 Procedure	45
2.5.3 Optimization of sequential injection determination of acetic acid	46
2.5.3.1 Effect of sodium hydroxide concentration	47
2.5.3.2 Effect of sodium hydroxide aspiration volume	50
2.5.3.3 Effect of phenolphthalein concentration	52
2.5.3.4 Effect of phenolphthalein aspiration volume	55
2.5.3.5 Effect of standard acetic acid aspiration volume	57
2.5.3.6 Effect of pump speed (flow rate)	59
2.5.3.7 Summary of conditions used	61
2.5.3.8 Precision	62
2.5.3.9 Determination of acetic acid in vinegar sample	64

	Page
CHAPTER 3 CONCLUSION	68
REFERENCES	69
APPENDIXES	71
APPENDIX A	72
APPENDIX B	73
APPENDIX C	74
APPENDIX D	79
APPENDIX E	82
APPENDIX F	85
APPENDIX G	88
APPENDIX H	89
CURRICULUM VITAE	90

LIST OF TABLES

Table	Page
2.1 List of vitamin C tablet samples	20
2.2 List of locally commercial vinegar samples	21
2.3 Preliminary conditions of ascorbic acid determination	25
2.4 Effect of acidic potassium permanganate concentration on peak area; mean of triplicate aspirations	26
2.5 Effect of acidic potassium permanganate concentration on linear equation and correlation coefficient	27
2.6 Effect of acidic potassium permanganate aspiration volume on peak area; mean of triplicate aspirations	28
2.7 Effect of acidic potassium permanganate aspiration volume on linear equation and correlation coefficient	29
2.8 Effect of ascorbic acid aspiration volume on peak area; mean of triplicate aspirations	31
2.9 Effect of acidic potassium permanganate concentration on linear equation and correlation coefficient	32
2.10 Effect of pump speed on peak height; mean of triplicate aspirations	32
2.11 Effect of pump speed on linear equation and correlation coefficient	33
2.12 The SIA conditions for ascorbic acid determination	34
2.13 Calibration range of standard ascorbic acid	34
2.14 A precision study	37
2.15 Calibration curve of ascorbic acid determination	39
2.16 Determination of ascorbic acid in vitamin C tablet samples	40
2.17 Calibration graph of SIA ascorbic acid	42
2.18 Recovery study	43
2.19 Study of percent recovery for SIA ascorbic acid determination	44
2.20 Preliminary conditions for acetic acid determination	47

Table	Page
2.21 Effect of sodium hydroxide concentration on peak area; mean of triplicate aspirations	48
2.22 Effect of sodium hydroxide concentration on linear equation and correlation coefficient	50
2.22 Effect of aspiration volume of sodium hydroxide solution on peak area; mean of triplicate aspirations	51
2.23 Effect of aspiration volume of sodium hydroxide solution on linear equation and correlation coefficient	52
2.25 Effect of phenolphthalein concentration on peak area; mean of triplicate aspirations	53
2.26 Effect of phenolphthalein concentration on linear equation and correlation coefficient	54
2.27 Effect of aspiration volume of phenolphthalein on peak area; mean of triplicate aspirations	56
2.28 Effect of aspiration volume of phenolphthalein on linear equation and correlation coefficient	57
2.29 Effect of aspiration volume of standard acetic acid solution on peak area; mean of triplicate aspirations	58
2.30 Effect of aspiration volume of standard acetic acid solution on linear equation and correlation coefficient	59
2.31 Effect of pump speed on peak area; mean of triplicate aspirations	60
2.32 Effect of pump speed on linear equation and correlation coefficient	61
2.33 The optimized SI condition for acetic acid determination.	62
2.34 Precision study using 5%(w/v) acetic acid and 4%(w/v) acetic acid (n=11)	63
2.35 Calibration graph of acetic acid	65
2.36 Determination of acetic acid in locally commercial vinegar sample	67

LIST OF FIGURES

Figure	Page
1.1 Typical sequential injection system utilizing programmed forward, reverse and stopped flow. HC, Holding coil: AW, auxiliary waste: and W, waste	2
1.2 Sequenced zones of reagent (R) and sample (S) intersperse during flow reversal while the detectable product (P) is formed. On arrival at the detector a suitable section of the product zone is captured within the observation field of the detector by selecting a stop delay time t_d for reaction-rate measurement	3
1.3 Principle of sequential injection technique showing manifold and readout. The plunger of the piston pump moves in discrete steps to aspirate wash solution, sample solution and reagent solution into a channel which consist of reaction (R), detection (D) and hold-up (H) sections. The selector valve provides sequential injection of the solutions into the channel through preprogrammed steps synchronized with piston movement. The record shows a typical double peak, the trough of which is located at the time when the flow has been reversed	4
1.4 Sequential injection manifolds designed to accommodate (A) two different assays, (B) sample and standard dilution, (C) analyte preconcentration and matrix removal and (D) gas diffusion and optosensing	6
1.4 (Continued) Sequential injection manifolds designed to accommodate (A) two different assays, (B) sample and standard dilution, (C) analyte preconcentration and matrix removal and (D) gas diffusion and optosensing	7
1.5 Concentration gradients, formed from sequentially injected zones of wash, sample and reagent in the flow channel at forward and reversed flow	8

Figure	Page
1.6 Sequential zone injection and their mutual penetration, as it take place during their forward flow in the channel (top). Individual (A, B) peaks of sample and reagent and their composite profile (A+B) show (below) the degree of mutual penetration and a region where A will react with B. Note that reversing of flow past t_3 will yield typical double peaks obtained in this work	10
1.7 Influence of repeated reversal (N) on mutual penetration of sample and reagent zone, obtained by injecting separately a non-reactive dye as a sample zone and then a reagent zone, and by performing multiple flow reversals in the auxiliary reactor. Since the dye zones entered the detector only during the final reversal period, only a single double peak is observed in all experiments. Note that whereas the residence time increases with the number of reversals, the peak height decreases and the peaks become progressively more symmetrical, yet the leading (sample) edge remains reagent free	11
1.8 A collagen Triple Helix	12
1.9 The structure of Vitamin C	13
1.10 Formation of ascorbic acid and its derivative	13
1.11 Distribution of FI determination of AA according to the detection technique	15
2.1 (a) A schematic diagram of the SIA system for ascorbic acid determination (b) A schematic diagram of order segment for ascorbic acid determination	23
2.2 SIA peaks for different acidic potassium permanganate concentration at a standard ascorbic acid concentration of 800 mg/L: (A) 0.001, (B) 0.0025 and (C) 0.005 M of acidic potassium permanganate	26

Figure	Page
2.3 Effect of acidic potassium permanganate concentration on peak area: (◆) 1.0×10^{-3} , (■) 2.5×10^{-3} , (▲) 5.0×10^{-3} M of acidic potassium permanganate	27
2.4 SIA peaks for different acidic potassium permanganate aspiration volume by using a standard ascorbic acid concentration of 800 mg/L: (A) 100, (B) 120, (C) 140, (D) 160, (E) 180, (F) 200 and (G) 220 μ L of acidic potassium permanganate	28
2.5 Effect of acidic potassium permanganate aspiration volume on peak area: (◆) 100, (■) 120, (▲) 140, (×) 160, (*), 180, (●) 200 and (⊕) 220 μ L of acidic potassium permanganate	29
2.6 SIA peaks for different ascorbic acid aspiration volume at a standard ascorbic acid concentration of 000 mg/L: (A) 120, (B) 100 and (C) 80 μ L of standard ascorbic acid	30
2.7 SIA peaks for different ascorbic acid aspiration volume at a standard ascorbic acid concentration of 000 mg/L: (◆) 80, (■) 100 and (▲) 120 μ L of a standard ascorbic acid solution	31
2.8 Effect of pump speed on peak height; mean of triplicate aspirations: (◆) 50, (■) 100, (▲) 150, (×) 200 and (*) 250 μ L/s of pump speed	33
2.9 Calibration range of standard ascorbic acid (up to 1600 mg/L): $y = -0.113x + 239.2, R^2 = 0.9890$	35
2.10 Calibration range of standard ascorbic acid (up to 1200 mg/L): $Y = -0.121 x + 242.3, R^2 = 0.9933$	35
2.11 SIA peaks for 11 replicates aspiration of 400 mg/L standard ascorbic acid solution	36
2.12 A precision study: (◆) 400 mg/L and (■) 1200 mg/L of standard ascorbic acid solution	38

Figure	Page
2.13 SIA peaks for calibration graph of ascorbic acid determination: (A) blank, (B) 400, (C) 800 and (D) 1200 mg/L of standard ascorbic acid	39
2.14 Calibration curve of ascorbic acid determination: $y = -0.133x + 317.0$, $R^2 = 0.9924$	40
2.15 SIA grams of a sample added with ascorbic acid standards: (A) Blank, (B) 400 and (C) 800 mg/L	41
2.16 Calibration graph ($y = 0.136x + 4.8$, $R^2 = 0.9903$)	42
2.17 Standard addition curve of SIA ascorbic acid determination, used (blank-area) data plotted (data of the sample code D): $y = 0.142x + 21.4$ $R^2 = 0.9997$	43
2.18 (a) A schematic diagram of the SIA system for acetic acid determination (b) A schematic diagram of the segment order for acetic acid determination	45
2.19 SIA peaks for different sodium hydroxide concentrations (a standard acetic acid concentration of 4%(w/v)): (A) 0.2, (B) 0.4, (C) 0.6, (D) 0.8 and 1.0 M of sodium hydroxide	48
2.20 Effect of sodium hydroxide concentration on peak area : (◆) 0.2, (■) 0.4, (▲) 0.6, (×) 0.8 and (*) 1.0 M of sodium hydroxide solution	49
2.21 SIA peaks for different aspiration volume of sodium hydroxide solutions (a standard acetic acid concentration of 4%(w/v)): (A) 50, (B) 60, (C) 70, (D) 80 and (E) 90 μ L of sodium hydroxide solution	50
2.22 Effect of aspiration volume of sodium hydroxide solution on peak area : (◆) 50, (■) 60, (▲) 70, (×) 80 and (*) 90 μ L of sodium hydroxide solution	51

Figure	Page
2.23 SIA peaks for different phenolphthalein concentration at a standard acetic acid concentration of 4%(w/v): (A) 8×10^{-3} , (B) 1×10^{-2} , (C) 2×10^{-2} , (D) 4×10^{-2} , (E) 6×10^{-2} and (F) 8×10^{-2}	53
2.24 Effect of phenolphthalein concentration on peak area: (◆) 8×10^{-3} , (■) 1×10^{-2} , (▲) 2×10^{-2} , (×) 4×10^{-2} , (✱) 6×10^{-2} and (●) 8×10^{-2} %(w/v) of phenolphthalein	54
2.25 SIA peaks for different aspiration volume of phenolphthalein (a standard acetic acid concentration of 4%(w/v)): (A) 50, (B) 60, (C) 70 and (D) 80 μL of 2×10^{-2} %(w/v) phenolphthalein	55
2.26 Effect of aspiration volume of phenolphthalein on peak : (×) 50, (◆) 60, (■) 70, (▲) 80 μL of 2×10^{-2} %(w/v) phenolphthalein	56
2.27 SIA peaks for different aspiration volume of standard acetic acid solution at a standard acetic acid concentration of 4%(w/v): (A) 70, (B) 60, (C) 50 and (D) 40 μL of standard acetic acid solution	58
2.28 Effect of aspiration volume of standard acetic acid solution on peak area : (■) 40, (▲) 50, (●) 60 and (✱) 70 μL of standard acetic acid solution	59
2.29 SIA peaks for different pump speed at a standard acetic acid concentration of 4% (w/v): (A) 200, (B) 175, (C) 150, (D) 125, (E) 100 and (F) 75 $\mu\text{L/s}$ of pump speed	60
2.30 Effect of pump speed on peak area : (◆) 75, (■) 100, (▲) 125, (×) 150, (✱) 175 and (●) 200 $\mu\text{L/s}$ of pump speed	61
2.31 SIA peaks for precision study using 5%(w/v) acetic acid (n=11)	62
2.32 Precision study (n=11): (◆) 5%(w/v) acetic acid and (■) 4%(w/v) of standard acetic acid	64
2.33 SIA peaks for calibration curve of acetic acid	65
2.34 Calibration curve of acetic acid: $y = -37.709x + 390.9$, $R^2 = 0.9989$	66
2.35 SIA peaks for vinegar sample of sample code A and E	66

ABBREVIATION AND SYMBOLS

SI	Sequential Injection
FI	Flow Injection
t_d	delay time
FI-MS	Flow Injection Mass Spectrometry
C_A^0	Initial concentration of A
C_B^0	Initial concentration of B
C_A	concentration of A
C_B	concentration of B
D_A	distribution coefficient of A
D_B	distribution coefficient of B
Abs	Absorbance
AOAC	American Official Associated Committee